

No. EN.DZ.SGE.25.U4

**EU Declaration of Conformity**

issued under the sole responsibility of the manufacturer

The object of this declaration: Hydrostatic level probes models **SGE-25, SGE-25S, SGE-25C**.Manufacturer: **APLISENS S.A.,  
ul. Morelowa 7, 03-192 Warszawa**

We hereby declare under the sole responsibility, that the object of the declaration defined above comply with relevant Union harmonization legislation.

Hydrostatic level probes models **SGE-25, SGE-25S, SGE-25C**

in all versions comply with directive:

- **EMC – 2014/30/EU** dated 26 February 2014

Conformity assessment procedure: module A. The following standards were applied: EN 61326-1:2013, EN 61326-2-3:2013.

- **RoHS – 2011/65/EU** dated 08 June 2011

Conformity assessment procedure: module A, according to Decision No 768/2008/EC of the European Parliament and of the Council.

The following standard was applied EN IEC 63000:2018.

Hydrostatic level probes models **SGE-25, SGE-25S, SGE-25C** in Ex versions comply with directive:

- **ATEX – 2014/34/EU** dated 26 February 2014

Level probes type SG-25 model ID 0039 0002 ....., SG-25S model ID 0040 0002 ....., SG-25C model ID 0041 0002 ..... in Intrinsically Safe version are marked with the following certificate marking:

I M1 Ex ia I Ma	I M1 Ex ia I Ma
II 1G Ex ia IIB T4/T5/T6 Ga lub	II 1G Ex ia IIC T4/T5/T6 Ga
KDB 09 ATEX 007X	KDB 09 ATEX 007X

Certificate (Module B) was issued by Notified Body no.1453, Central Mining Institute, Plac Gwarków 1, 40-166 Katowice.

The following standards were applied: EN IEC 60079-0:2018, EN 60079-11:2012, EN 50303:2000.

Level probes type SG-25 model ID 0039 0004 ....., SG-25S model ID 0040 0004 ....., SG-25C model ID 0041 0004 ..... in Intrinsically Safe version are marked with the following certificate marking:

I M1 Ex ia I Ma	I M1 Ex ia I Ma	I M1 Ex ia I Ma
II 1G Ex ia IIB T6/T5/T4 Ga lub	II 1G Ex ia IIC T6/T5/T4 Ga lub	II 2G Ex ia IIC T6/T5/T4 Gb
FTZU 12 ATEX 0193X	FTZU 12 ATEX 0193X	FTZU 12 ATEX 0193X

Conformity assessment procedure: module B. Notified Body no.1026, Physical Technical Testing Institute, Pikartská 1337/7, 716 07 Ostrava Radvanice.

The following standards were applied: EN IEC 60079-0:2018, EN 60079-11:2012, EN 50303:2000.

Notification of quality assurance is issued (Module D) by Notified Body no.1453, Central Mining Institute, Plac Gwarków 1, 40-166 Katowice.

Warsaw, 22.05.2023

Adam Żurawski  
General Manager